


IMPLANT ONE **ABUTMENT SCREW**

SPECIFICATIONS
Material: 6AL4V E.L.I. Titanium
Diameter: 1.1 mm – 1.4 mm
Height: 9.6 mm
.050" hex interface



The tip of the abutment screw should protrude approximately 2.6mm (5 to 6 exposed threads).

When turning the screw into the implant, the screw should turn 5 to 6 turns for proper thread engagement. Do not exceed specified torque.

